

CALCULATE RESULTS

1. Semi-quantitative results can be derived by simple comparison of the sample absorbances to the absorbance of the calibrator wells. Samples containing less color than a calibrator well have a concentration of Cyndrospermopsin greater than the concentration of the calibrator. Samples containing more color than a calibrator well have a concentration of Cyndrospermopsin less than the concentration of the calibrator.

2. Quantitative interpretation requires graphing the absorbances of the calibrators (Y axis) versus the log of the calibrator concentration (X axis) on semi-log graph paper. A straight line is drawn through the calibrator points and the sample absorbances are located on the line. The corresponding point on the X-axis is the concentration of the sample. Samples with absorbances greater than the lowest calibrator or less than the highest calibrator must be reported as < 0.1 ppb or >2 ppb, respectively. Alternatively, Beacon can supply a spreadsheet template that can be used for data reduction. Please contact Beacon for further details.

SAMPLE CALCULATIONS

Well Contents	OD	Average OD ± SD*	%RSD	%Bo**	Cyldrosper- mopsin Conc. (ppm)
Negative Control	1.487 1.444	1.466 ± 0.030	2.07	100	N/A
0.1ppb Calibrator	1.153 1.155	1.154 ± 0.001	0.12	78.7	N/A
0.5 ppb Calibrator	0.699 0.703	.701 ± 0.003	0.4	47.8	N/A
2 ppb Calibrator	0.268 0.261	0.265 ± 0.005	1.87	18	N/A
Sample	0.487 0.497	0.492 ± 0.006	1.44	33.6	0.9355

Actual values may vary; this data is for example purposes only.

* standard deviation

** %Bo equals average sample absorbance divided by average negative control absorbance times 100%.

TECHNICAL ASSISTANCE

For questions regarding this kit or for additional information about Beacon products, call (207) 571-4302.

SAFETY

To receive complete safety information on this product, contact Beacon Analytical Systems, Inc. and request Material Safety Data Sheets. Stop Solution is 1N hydrochloric acid. Handle with care.

General Limited Warranty

Beacon Analytical Systems, Inc. ("Beacon") warrants the products manufactured by it against defects in materials and workmanship when used in accordance with the applicable instructions for a period not to extend beyond a product's printed expiration date. BEACON MAKES NO OTHER WARRANTY, EXPRESSED OR IMPLIED. THERE IS NO WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. The warranty provided herein and the data, specifications and descriptions of Beacon products appearing in published catalogues and product literature may not be altered except by express written agreement signed by an officer of Beacon. Representations, oral or written, which are inconsistent with this warranty or such publications are not authorized and, if given, should not be relied upon.

In the event of a breach of the foregoing warranty, Beacon's sole obligation shall be to repair or replace, at its option, any product or part thereof that proves defective in materials or workmanship within the warranty period, provided the customer notifies Beacon promptly of any such defect. The exclusive remedy provided herein shall not be deemed to have failed of its essential purpose so long as Beacon is willing and able to repair or replace any nonconforming Beacon product or part. Beacon shall not be liable for consequential, incidental, special or any other indirect damages resulting from economic loss or property damage sustained by a customer from the use of its products. However, in some states the purchaser may have rights under state law in addition to those provided by this warranty.



Cyldrospermopsin Plate Kit

Cat. # 20-0149

Instructional Booklet

READ COMPLETELY BEFORE USE.

BEACON ANALYTICAL SYSTEMS, INC.

82 Industrial Park Road

Saco, ME 04072

Tel. (207) 571-4302

Fax (207) 602-6502

www.beaconkits.com

INTENDED USE

The Beacon Cyldrospermopsin Plate Kit is a competitive ELISA for the quantitative analysis of Cyldrospermopsin in water.

USE PRINCIPLES

The Beacon Cylindrospermopsin plate kit is a competitive enzyme-labeled immunoassay. The Cylindrospermopsin -HRP enzyme conjugate is pipetted into test wells followed by calibrators or sample extracts. Cylindrospermopsin Antibody Solution is then pipetted into the test wells to initiate the reaction. During the 45-minute incubation period, Cylindrospermopsin from the sample and Cylindrospermopsin HRP conjugate compete for binding to Cylindrospermopsin antibody. The Cylindrospermopsin antibody is captured on the walls of the test well. Following this 45-minute incubation, the contents of the well are removed and the wells are washed to remove any unbound Cylindrospermopsin, Cylindrospermopsin HRP conjugate and free Cylindrospermopsin antibody. After this wash step, a clear substrate is then added to the wells and any bound enzyme conjugate causes the conversion to a blue color. Following a 45-minute incubation, the reaction is stopped and the amount of color in each well is read. The color of the unknown samples is compared to the color of the calibrators and the Cylindrospermopsin concentration of the samples is derived.

MATERIALS PROVIDED IN THE BEACON CYLINDROSPERMOPSIN PLATE KIT

The kit in its original packaging can be used until the end of the month indicated on the box label when stored at 2 – 8°C.

- 1 plate containing 12 test strips of 8 wells each vacuum-packed in aluminized pouch with indicating desiccant.
- 4 vials each containing 2 mL of Cylindrospermopsin calibrators corresponding to 0, 0.1, 0.5, 2 µg/L (ppb) of Cylindrospermopsin.
- 1 vial containing 7 mL Cylindrospermopsin HRP Enzyme Conjugate.
- 1 vial containing 7 mL of Polyclonal anti-Cylindrospermopsin antibody.
- 1 vial containing 14 mL of Substrate.
- 1 vial containing 14 mL of Stop Solution. (Caution! 1N HCl. Handle with care.)
- 1 Instructional Booklet

PERFORMANCE CHARACTERISTICS

SPECIFICITY

The Beacon Cylindrospermopsin cannot differentiate between Cylindrospermopsin and related compounds, but detects their presence to differing degrees.

COMPOUND	% CR
CYLINDROSPERMOPSIN	100 %
MICROCYSTIN-LR	< 1%
NODULARIN	< 1%

MATERIALS REQUIRED BUT NOT PROVIDED

- Methanol
- Pipet with disposable tips capable of dispensing 50 µL.
- Multi-channel pipet; 8 channel capable of dispensing 50 and 100 µL.
- Paper towels or equivalent absorbent material.
- Microwell plate or strip reader with 450nm filter.
- Timer
- Vortex mixer
- Wash bottle
- Laboratory quality distilled or deionized water.
- Graduated cylinder, 100 ml or larger.
- Glassware for sample collection and dilution.

PRECAUTIONS

1. Each reagent is optimized for use in the Beacon Cylindrospermopsin Plate Kit. Do not substitute reagents from any other manufacturer into the test kit. Do not combine reagents from other Beacon Cylindrospermopsin Plate Kits with different Lot numbers.
2. Dilution or adulteration of reagents or samples not called for in the procedure may result in inaccurate results.
3. Do not use reagents after expiration date.
4. Reagents should be brought to room temperature, 20 – 28°C (62 – 82°F) prior to use. Avoid prolonged (> 24 hours) storage at room temperature.
5. Cylindrospermopsin is a toxin and should be treated with care.
6. The Stop Solution is 1N hydrochloric acid. Avoid contact with skin and mucous membranes. Immediately clean up any spills and wash area with copious amounts of water. If contact should occur, immediately flush with copious amounts of water.

ASSAY PROCEDURE

(Note: Running calibrators and samples in duplicate will improve assay precision and accuracy.)

1. Allow reagents and sample extracts to reach room temperature prior to running the test. Fill a wash bottle with lab grade water.
 2. Place the appropriate number of test wells into a micro well holder. Be sure to re-seal unused wells in the zip-lock bag with desiccant.
 3. Using a pipet with disposable tips, add **50 µL of Enzyme conjugate** to each well.
 4. Add **50 µL of Calibrators or sample extract** to the appropriate test wells. Be sure to use a clean pipet tip for each.
 5. Dispense **50 µL of Antibody Solution** into each test well.
 6. Shake the plate gently for 30 seconds and incubate the test wells for **45 minutes**.
 7. Dump the contents of the wells into an appropriate waste container. Fill the wells to overflowing with laboratory grade water and dump. Repeat 3X for a total of four washes.
 8. Following the last wash tap the inverted wells onto absorbent paper to remove the last of the wash solution.
 9. Dispense **100 µL of Substrate** into each well.
 10. Incubate the wells for **45 minutes**.
 11. Dispense **100 µL of Stop Solution** into each test well.
 12. Read and record the absorbance of the wells at 450 nm using a strip or plate reader.
-